


UNDERWATER QUANTUM (PAR) METERS

MQ-510 & MQ-210

Spectral Response

Research-grade measurements of underwater photosynthetically active radiation






Mean **spectral response** measurements of six replicate Apogee MQ-210 and MQ-510 series quantum sensors.

Product Specifications

	MQ-510 MQ-210				
Calibration Uncertainty	± 5 %				
Measurement Range	0 to 4000 μmol m ⁻² s ⁻¹				
Measurement Repeatability	Less than 0.5 %				
Long-term Drift (Non-stability)	Less than 2 % per year				
Non-linearity	Less than 1 % (up to 4000 μ mol m $^{-2}$ s $^{-1}$)				
Response Time	Less than 1 ms				
Field of View	180°				
Spectral Range	389 to 692 nm ± 5 nm (wavelengths where response is greater than 50 % of maximum)	410 to 655 nm (wavelengths where response is greater than 50 % of maximum)			
Spectral Selectivity	Less than 10 % from 412 to 682 nm ± 5 nm	Less than 10 % from 469 to 655 nm \pm 5 nm			
Directional (Cosine) Response	± 5 % at 75° zenith angle				
Temperature Response	-0.11 ± 0.04 % per C	0.06 ± 0.06 % per C			
Uncertainty in Daily Total	Less than 5 %				
Detector	Blue-enhanced silicon photodiode				
Housing	Anodized aluminum body with acrylic diffuser				
IP Rating	IP68				
Operating Environment	0 to 50 C; less than 90 % non-condensing relative humidity up to 30 C; less than 70 % non-condensing relativity humidity from 30 to 50 C; separate sensors can be submerged in water up to depth of 30 m				
Meter Dimensions	126 mm length, 70 mm width, 24 mm depth	114 mm length, 60 mm width, 14 mm depth			
Sensor Dimensions	24 mm diameter, 37 mm height	24 mm diameter, 33 mm height			
Mass	180 g				
Cable	2 m of shielded, twisted-pair wire; additional cable available; TPR jacket				
Warranty	4 years against defects in materials and workmanship				

Dimensions



Spectral Errors

Features

DESIGNED FOR UNDERWATER USE

Sensor heads are fully epoxy potted to be completely waterproof. Diffuser is cosine corrected for accurate 2-pi PAR-mapping. Sensor readings are adjusted in firmware to correct for the immersion effect.

ACCURATE, STABLE MEASUREMENTS

Long-term non-stability determined from multiple replicate quantum sensors in accelerated aging tests and field conditions is less than 2 % per year.

DATALOGGING CAPABILITIES

The meter records up to 99 measurements in logging mode, making automatic measurements every 30 seconds and recording 30-minute averages. Data can be downloaded to calculate DLI.

	Apogee SQ-500	Apogee SQ-110 SQ-120	LI-COR LI-190	Kipp & Zonen PQS 1
Sun (Clear Sky)	0.0	0.0	-0.4	-1.0
Sun (Cloudy Sky)	0.1	0.2	-0.2	-1.3
Sun (Reflected from Grass Canopy)	-0.3	3.8	-0.8	1.1
Sun (Transmitted below Wheat Canopy)	0.1	4.5	-0.1	-0.3
Cool White Fluorescent (T5)	0.0	0.0	0.0	0.0
Metal Halide	0.9	-2.8	0.2	-1.7
Ceramic Metal Halide	0.3	-16.1	0.4	-0.7
High Pressure Sodium	0.1	0.2	1.3	1.4
Red LED (667 nm peak, 20 nm full-width half-maximum)	2.8	-62.1	3.5	-1.8
Red, Blue, White LED Mixture (60 % Red, 25 % White, 15 % Blue)	-2.0	-35.5	2.6	-1.7

NIST TRACEABLE CALIBRATION

Apogee Quantum sensors are calibrated by comparison to the mean of four transfer standard sensors under a reference lamp. The reference sensors are recalibrated regularly to a halogen lamp traceable to the National Institute of Standards and Technology. Calibration certificates are available upon request.

