

UNDERWATER PAR/ePAR METERS


MQ-650, MQ-510, & MQ-210X



Spectral Responses

Research-grade measurements of underwater photosynthetically active radiation






Mean **spectral response** measurements of six replicate Apogee MQ-210X (black), MQ-650 (green), and MQ-510 (blue) series quantum sensors compared to traditionally defined PAR 400-700 nm (dotted).

Product Specifications

	MQ-650	MQ-510	MQ-210X			
Calibration Uncertainty	± 5 %					
Measurement Range	0 to 4000 μmol m ⁻² s ⁻¹					
Measurement Repeatability	Less than 0.5 %					
Long-term Drift (Non-stability)	Less than 2 % per year					
Non-linearity	Less than 1 % (up to 4000 μ mol m ⁻² s ⁻¹)					
Response Time	Less than 1 ms					
Field of View	180°					
Spectral Range	400 to 750 nm ± 5 nm	389 to 692 nm ± 5 nm	370 to 650 nm ± 5 nm			
Directional (Cosine) Response	± 5 % at 75° zenith angle					
Temperature Response	-0.11 ± 0.	-0.04 % per C				
Uncertainty in Daily Total	Less than 5 %					
Detector	Blue-enhanced silicon photodiode					
Housing	Anodized aluminum body with acrylic diffuser					
IP Rating	IP68					
Operating Environment	0 to 50 C; less than 90 % non-condensing relative humidity up to 30 C; less than 70 % non-condensing relativity humidity from 30 to 50 C; separate sensors can be submerged in water up to depth of 30 m					
Meter Dimensions	126 mm length, 70 mm width, 24 mm depth					
Sensor Dimensions	30.5 mm diameter, 37 mm height	24 mm diameter, 37 mm height	24 mm diameter, 33 mm height			
Mass	180 g					
Cable	2 m of shielded, twisted-pair wire; additional cable available; TPR jacket					
Warranty	4 years against defects in materials and workmanship					



Features

DESIGNED FOR UNDERWATER USE

Sensor heads are fully epoxy potted to be completely waterproof. Diffuser is cosine corrected for accurate 2-pi PAR-mapping. Sensor readings are adjusted in firmware to correct for the immersion effect.

ACCURATE, STABLE MEASUREMENTS

Long-term non-stability determined from multiple replicate quantum sensors in accelerated aging tests and field conditions is less than 2 % per year.

DATALOGGING CAPABILITIES

The meter records up to 99 measurements in logging mode, making automatic measurements every 30 seconds and recording 30-minute averages. Data can be downloaded to calculate DLI.

	Apogee SQ-500	Apogee SQ-210X	LI-COR LI-190	Kipp & Zonen PQS 1
Sun (Clear Sky)	0.0	0.0	-0.4	-1.0
Sun (Cloudy Sky)	0.1	0.2	-0.2	-1.3
Sun (Reflected from Grass Canopy)	-0.3	5.0	-0.8	1.1
Sun (Transmitted below Wheat Canopy)	0.1	7.0	-0.1	-0.3
Cool White Fluorescent (T5)	0.0	7.2	0.0	0.0
Metal Halide	0.9	6.9	0.2	-1.7
Ceramic Metal Halide	0.3	-0.9	0.4	-0.7
High Pressure Sodium	0.1	3.2	1.3	1.4
Red LED (667 nm peak, 20 nm full-width half-maximum)	2.8	-30.9	3.5	-1.8
Red, Blue, White LED Mixture (60 % Red, 25 % White, 15 % Blue)	-2.0	-21.2	2.6	-1.7

NIST TRACEABLE CALIBRATION

Apogee Quantum sensors are calibrated by comparison to the mean of four transfer standard sensors under a reference lamp. The reference sensors are recalibrated regularly to a halogen lamp traceable to the National Institute of Standards and Technology. Calibration certificates are available upon request.



Spectral Errors