

# Specifications

|                       | QSOA-S (sunlight)   | QSOA-E (electric)  |
|-----------------------|---|--|
| Calibration           | Natural sunlight  | Electric lamps (8% difference)   |
| Absolute accuracy     | ± 5 %   | ± 5 %  |
| Uniformity            | ± 3 %   | ± 3 %  |
| Repeatability         | ± 1 %   | ± 1 %  |
| Application           | Measuring Photosynthetic Photon Flux  |  |
| 2.5 V option          | Output: 0 to 2.5 V (2.0 V = full sunlight 2,000 $\mu\text{mol m}^{-2} \text{s}^{-1}$ )  |  |
|                       | Input power   | 2.5 to 5.5 VDC   |
|                       | Sensitivity   | Custom calibrated to exactly 1.00 $\mu\text{mol m}^{-2} \text{s}^{-1} / \text{mV}$ |
| 5.0 V option          | Output: 0 to 5 V (4.0 V = full sunlight 2,000 $\mu\text{mol m}^{-2} \text{s}^{-1}$ )  |  |
|                       | Input power   | 5 to 5.5 VDC   |
|                       | Sensitivity   | Custom calibrated to exactly 0.50 $\mu\text{mol m}^{-2} \text{s}^{-1} / \text{mV}$ |
| Current draw          | 285 $\mu\text{A}$   |  |
| Operating environment | - 40 to 55 °C; 0 to 100% relative humidity. Designed for continuous outdoor use. Can be submerged under water (with or without mounting screw). |  |
| Materials             | Anodized aluminum with acrylic lens   |  |
| Cable                 | 3 meters of shielded, twisted-pair wire with Santoprene casing, ending in pigtail leads. Additional cable \$1.95/meter.                         |  |
| Dimensions            | 2.4 cm diameter, 2.75 cm high   |  |
| Mass                  | 70 g (with 3 m lead wire)   |  |
| Warranty              | 1 year against defects in materials and workmanship   |  |



435-792-4700

✉ [www.apogeeinstruments.com](http://www.apogeeinstruments.com)  
[techsupport@apogee-inst.com](mailto:techsupport@apogee-inst.com)

## Amplified Quantum Sensor

### Owners Manual

Model: QSOA-E  
QSOA-S

2.5 and 5.0 V options


apogee  
Logan UT

## Calibration

Quantum sensors are calibrated for electric light or sunlight. Average spectral errors associated with each calibration are shown below.

|  |                        | Electric Calibration | Sunlight Calibration |
|--|------------------------|----------------------|----------------------|
|  | Cool White Fluorescent | 0% error             | 8% high              |
|  | Metal Halide           | 0% error             | 8% high              |
|  | High Pressure Sodium   | 6% low               | 2% high              |
|  | Sunlight               | 8% low               | 0% error             |

## Setup Instructions



**Red:** positive (signal from sensor)

**Black:** input power

**Clear:** ground (for sensor signal and input power)


|                   | 2.5 option  | 5.0 option  |
|-------------------|---|---|
| Power Supply      | 2.5 to 5.5 V  | 5.0 to 5.5 V  |
| Conversion factor | 1.0 $\mu\text{mol m}^{-2} \text{s}^{-1}$ per mV     | 0.5 $\mu\text{mol m}^{-2} \text{s}^{-1}$ per mV     |
| Output (volts)    | 0.0 to 2.5 V  | 0.0 to 5.0 V  |
| Full sunlight     | 2.0 V (2,000 $\mu\text{mol m}^{-2} \text{s}^{-1}$ ) | 4.0 V (2,000 $\mu\text{mol m}^{-2} \text{s}^{-1}$ ) |

**Do not exceed 5 Volts in power supply.**

# Quantum Sensor Models

A quantum refers to the amount of energy carried by a photon. Apogee quantum sensors approximate the quantity of photons between 400 and 700 nanometers. Photosynthesis is largely driven by the number of photons between these wavelengths, so this radiation is called the Photosynthetic Photon Flux (PPF) and is measured in  $\mu\text{mol m}^{-2} \text{s}^{-1}$ .

## Amplified Quantum Sensors:



The model, serial number, production date, and calibration factor are located on the sensor cable.



2

### Cosine response

Some of the radiation coming into a sensor at low angles is reflected, which causes low readings. The convex optical disc is designed to capture radiation at low angles and minimize cosine response errors. The cosine error for typical applications is less than 2%.

### Temperature response


The temperature response is less than 0.1% per degree Celsius. This temperature error is not significant in most applications.

### Long-term stability

Our research indicates that the average output increases approximately 1% per year because of changes in the optical transparency of the diffusion disk. We recommend returning the sensor for recalibration every 2 years.

### Spectral Response

As shown in the graph below, quantum response by definition is from 400 to 700 nm, and gives equal emphasis to all photons in that range. The spectral response of the Apogee sensor, as well as a typical plant response, are also shown.




7

## Mounting the QSOA-E and QSOA-S




Each sensor is equipped with a mounting bolt. Mount the sensor as level as possible. Small changes in level can cause measurement errors. We recommend using our leveling plate (model LEV) for the most accurate measurements. The sensor should be mounted with the cable pointing toward the nearest magnetic pole to minimize azimuth error.



4

## Calibration



PPF = sensor output x conversion factor  
 =  $4,000 \text{ mV} \times 0.5 \mu\text{mol m}^{-2} \text{s}^{-1} \text{ per mV} = 2,000 \mu\text{mol m}^{-2} \text{s}^{-1}$

5