

721 W 1800 N Logan, UT 84321

Certificate of Calibration

Reference Pyranometer $[V_R]$: Test Pyranometer $[V_T]$:

Manufacturer: Kipp & Zonen Manufacturer: Kipp & Zonen

Model:CMP 10Model:EXAMPLESerial Number:140713Serial Number:EXAMPLE

Responsivity [R_R]: 9.351 μ V/Wm⁻² Responsivity [R_T]: 9.647 μ V/Wm⁻²

Initial Position:

V_{T111} = V_{R111} = 4.630 m۷ 4.350 m۷ $V_{T1S} =$ $V_{R1S} =$ -0.010 m۷ 0.000 m۷ $V_{R1} =$ 4.640 m۷ 4.350 m۷

Transposed Position:

 $V_{R2U} =$ $V_{T2U} =$ 4.790 4.210 m۷ m۷ $V_{R2S} =$ 0.000 $V_{T2S} =$ 0.010 m۷ m۷ $V_{R2} =$ 4.210 $V_{T2} =$ 4.780 m۷

Calibration Stability:

 $(0.995) < \frac{V_{R1} V_{T1}}{V_{R2} V_{T2}} < (1.005)$ 1.003 Pass

Test Pyranometer Responsivity:

 $R_T = \frac{V_{R1} + V_{T1}}{V_{R2} + V_{T2}} R_R$ 9.647 $\mu V/Wm^{-2}$

Calibration Factor: 103.662 Wm⁻²/mV

Calibration Procedure

Calibration is based on a side-by-side comparison under high intensity metal halide lamps using an ISO-classified reference pyranometer of the same type, and is in accordance with ISO 9847 type IIc. The ISO-classified reference pyranometer is recalibrated on an annual schedule under clear sky conditions in Logan, Utah using the mean of at least (2) ISO-9060 Spectrally Flat Class A reference pyranometers, and is in accordance with ISO 9847 type Ia. Each of the four Class A reference pyranometers are recalibrated on an alternating year schedule (two instruments per year) at the National Renewable Energy Laboratory (NREL) in Golden, Colorado. NREL reference standards are calibrated to the World Radiometric Reference (WRR) in Davos, Switzerland.

Traceability		
Reference Instrument	Serial Number	ISO 9060 Classification
EKO Instruments MS80*	S16088044	Spectrally Flat Class A
Kipp & Zonen CM11	060089	Spectrally Flat Class A
Kipp & Zonen CMP11	101625	Spectrally Flat Class A
Hukseflux SR20	2497	Spectrally Flat Class A
Kipp & Zonen CMP10	140713	Spectrally Flat Class A

Technical Manager: Jacob Birgham Date: 13-Jan-2020